


BUTTER TRIGLYCERIDES C24 – C56



Column

Phase	MEGA-1
I.D.	0.32 mm
Film Thickness	0.1 µm
Length	3 m + 1 m Retention Gap

Chromatographic Conditions

Injection	On column	
Injection volume	1 µL	
Oven	T start	60°C
	Rate	15°C/min
	T end	340°C
Detector	FID	350°C
Carrier Gas	Hydrogen	30 kPa

Peak Identification

1	C 24
2	Cholesterol
3	C 26
4	C 28
5	C 30
6	C 32
7	C 34
8	C 36
9	C 38
10	C 40
11	C 42
12	C 44
13	C 46
14	C 48
15	C 50
16	C 52
17	C 54
18	C 56

