

**PESTICIDES**

Courtesy of Prof. C. Bicchi, C. Brunelli – Università di Torino, Dipartimento Scienza e Tecnologia del Farmaco – Via P.Giuria, 9 – Torino

**Column**

Phase	MEGA-1701
I.D.	0.1 mm
Film Thickness	0.1 µm
Length	5 m

**Chromatographic Conditions**

Injector	Split	230°C
Injection Volume	1.0 µL	
Sample Dilution	1:200	in Cyclohexane
Oven	T start	50°C (0.1 min)
	Rate	50°C/min
	T end	250°C (5 min)
Detector	FID	250°C
Carrier Gas	Hydrogen	0.5 mL/min

**Peak Identification**

1	α-HCH
2	γ-HCH
3	Heptachlor
4	Chlorotalonil
5	/
6	Parathion-Me
7	Malathion
8	Fenotrothion
9	Parathion-Et
10	/
11	Fenitrothion
12	Chlordane-Cis + Trans
13	Dieldrin
14	o,p'-DDT
15	β-Endosulfan
16	p,p'-DDT
17	/
18	Tetradifon

