


PHTHALATES – EPA METHOD 606



Column

Phase	MEGA-1
I.D.	0.53 mm
Film Thickness	1.5 µm
Length	15 m

Chromatographic Conditions

Injector	Split	250°C
Injection Volume	1.0 µL	
Oven	T start	150°C
	Rate	15°C/min
	T end	270°C
Detector	FID	300°C
Carrier Gas	Helium	20 mL/min

Peak Identification

1	Dimethyl phthalate
2	Diethyl phthalate
3	Di-n-butyl phthalate
4	Butyl Benzyl phthalate
5	Bis (2-ethyhexyl) phthalate
6	Di-n-octyl phthalate