

Phthalate Esters EPA Method 606

Column: **MEGA-1 - 0.53mm, 1.50 μ m, 15m**

Catalog Code: S-1-053-150-15

Conditions:

Injection: Split 250°C, 1.0 μ L.

Sample: 2000 μ g/mL each in Methanol.

Detector: FID 300°C.

Oven Temp.: 150°C, 15°C/min, 270°C.

Carrier Gas: Helium, 20mL/min.

Peaks:

1. Dimethyl phthalate
2. Diethyl phthalate
3. Di-n-butyl phthalate
4. Butyl benzyl phthalate
5. Bis (2-ethylhexyl) phthalate
6. Di-n-octyl phthalate

